



---

## **Collusion and falsification of results in proficiency testing**

Asia Pacific Food Analysis Network (APFAN)  
APFAN PT2 Workshop  
"Food Analysis: Proficiency Testing and Reference Materials",  
Bangkok, Thailand  
19<sup>th</sup>-21<sup>st</sup> June 2019.



---

PROFICIENCY TESTING AUSTRALIA

**Philip Briggs**  
**General Manager, PTA**  
**Associate Member, APLAC**  
**Executive Member, ICASI**

## Overview

---

- **ISO/IEC 17043 technical requirements**
- **Collusion**
- **Falsification of results**
- **Detection of suspect results**
- **Program design and actions**

## Introduction

---

**The main purpose proficiency testing is to provide an evaluation of the performance of laboratories for specific tests or measurements and to monitor laboratories' continuing performance.**

## Introduction

---

**PT participation - in a short time by multiple laboratories working in the same scientific field.**

**Common samples received in the form of a 'test' may lead to a change in laboratory operations.**

## Introduction

---

**To provide a true evaluation of performance then the design of the programs should be to encourage the laboratories to treat the PT samples as **routine samples** received from a commercial client.**

## ISO/IEC 17043 Requirements

---



**Section 4.4.1.3 j) states - “reasonable precautions to prevent collusion between participants or falsification of results, and procedures to be employed if collusion or falsification of results is suspected.”**

## ISO/IEC 17043 Requirements

---



**This requirement may be considered to place an unfair load of responsibility on the PT provider to attempt control the ethical and professional behaviour of organisations that have registered to participate in their PT programs.**

## Collusion

---



**Collusion** in PT may mean a form of cheating from the participating laboratory. Example - two or more people from two or more laboratories work together in a deceitful way to submit results for their benefit.

## Collusion

---



**Multiple staff involved from one laboratory are involved to submit results – is this collusion?**

## Collusion

---



**Multiple laboratories from the same organisation are involved to submit results – is this collusion?**

## Collusion

---



**Multiple laboratories from different organisations are involved to submit results – is this collusion?**

## Falsification

---



**Falsification** of results in PT may mean the submission of results are stated untruthfully or are being misrepresented as being from the participating laboratory.

## Falsification

---



**PT accepts results are from the identified laboratory in “good faith” –**

- **Laboratory code**
- **Signature**
- **Laboratory logo**

## Falsification

**If the falsification of results is due to collusion then this may be detected quite easily if the results are identical.**

## Falsification

Aluminium							
Results by Laboratory Code							
Laboratory Code	Sample PTA 2A						
	Result	±	MU <sup>1</sup>	Robust z-score <sup>2</sup>	Method Code <sup>3</sup>	Digestion Code <sup>3</sup>	
	mg/L						
103	4.05	±	0.10	0.16	2	14	
129	3.83	±	0.113	-1.58	2	#	
181	3.84		#	-1.50	2	7	
230	3.71	±	0.0228	-2.53	2	7	
233	4.04	±	0.063	0.08	2	14	
303	3.99	±	0.20	-0.32	2	14	
345	3.71	±	0.13	-2.53	2	14	
359	4.04	±	0.063	0.08	2	14	
369	4.05	±	0.06	0.16	2	14	



## Falsification

**If the results reported were similar to other participants reported results then the falsification of results in this case would be quite difficult or impossible to detect.**

## Falsification

Aluminium							
Results by Laboratory Code							
Laboratory Code	Sample PTA 2A						
	Result	± mg/L	MU <sup>1</sup>	Robust z-score <sup>2</sup>	Method Code <sup>3</sup>	Digestion Code <sup>3</sup>	
103	4.05	±	0.10	0.16	2	14	
129	3.83	±	0.113	-1.58	2	#	
181	3.84		#	-1.50	2	7	
230	3.71	±	0.0228	-2.53	2	7	
233	4.04	±	0.063	0.08	2	14	
303	3.99	±	0.20	-0.32	2	14	
345	3.71	±	0.13	-2.53	2	14	
359	4.074	±	0.815	0.35	2	14	
369	4.05	±	0.06	0.16	2	14	

## Falsification

The participating laboratory sub-contracts the testing to another laboratory and then submit these results as their own to the proficiency testing provider.

Can the PT provider detect?

## Falsification

Barium						
Results by Laboratory Code						
Laboratory Code	Sample PTA 1B					
	Result	±	MU <sup>1</sup>	Robust z-score <sup>2</sup>	Method Code <sup>3</sup>	Digestion Code <sup>3</sup>
	mg/L					
103	0.092	±	0.002	-0.17	2	14
129	0.090	±	0.023	-0.51	2	#
181	0.104		#	1.85	2	7
230	0.098	±	0.0176	0.84	2	7
233	0.092	±	0.051	-0.17	2	14
303	0.0934	±	0.005	0.07	2	14
345	<0.2		#	na	2	14
359	0.090	±	0.009	-0.51	2	14
369	0.0944	±	0.0017	0.24	2	14
381	0.084	±	0.014	-1.52	3	14

## Detection of suspect results

**Identical results reported by two or more laboratories.**

## Detection of suspect results

Boron							
Results by Laboratory Code							
Laboratory Code	Sample PTA 1A						
	Result	± mg/L	MU <sup>1</sup>	Robust z-score <sup>2</sup>	Method Code <sup>3</sup>	Digestion Code <sup>3</sup>	
103	0.080	±	0.011	0.47	2	14	
129	0.0783	±	0.0009	0.31	2	14	
181	0.050		#	-2.33	2	7	
230	0.083	±	0.0282	0.74	2	7	
233	0.141	±	0.069	6.14	§ 2	14	
303	0.0783	±	0.0009	0.31	2	14	
345	0.119	±	0.0119	4.09	§ 2	14	
359	0.065	±	0.011	-0.93	3	14	
369	0.0783	±	0.0009	0.31	2	14	

## Detection of suspect results

**Narrow range of expected values  
then high probability that the results  
reported may be identical.**

## Detection of suspect results

Bismuth							
Results by Laboratory Code							
Laboratory Code	Sample PTA 2A						
	Result	±	MU <sup>1</sup>	Robust z-score <sup>2</sup>	Method Code <sup>3</sup>	Digestion Code <sup>3</sup>	
	mg/L						
103	4.13	±	0.29	0.40	3	14	
129	4.13	±	0.088	0.40	2	#	
181	3.99		#	-0.46	2	7	
230	4.12	±	0.0172	0.34	2	7	
345	4.12	±	0.161	0.34	2	14	
359	4.167	±	0.625	0.63	2	14	
362	7.18	±	1.00	19.10	§	2	14
369	4.09	±	0.10	0.16	2	14	
381	4.09	±	0.10	-2.97	3	14	



## Detection of suspect results

**The probability of identical results being reported as coincidence lessens when the reporting accuracy increases and lessens further upon replicate reporting for single or multiple samples.**



## Detection of suspect results

Boron							
Results by Laboratory Code							
Laboratory Code	Sample PTA 1A						
	Result	± mg/L	MU <sup>1</sup>	Robust z-score <sup>2</sup>	Method Code <sup>3</sup>	Digestion Code <sup>3</sup>	
103	0.0783	±	0.0009	0.31	2	14	
129	0.069	±	0.030	-0.56	2	#	
181	0.050		#	-2.33	2	7	
230	0.083	±	0.0282	0.74	2	7	
233	0.141	±	0.069	6.14	§	2	14
303	0.075	±	0.004	0.00	2	14	
345	0.119	±	0.0119	4.09	§	2	14
359	0.065	±	0.011	-0.93	3	14	
369	0.0783	±	0.0009	0.31	2	14	

## Program design

---

**Programs with large participation may be divided into two or more groups with each group receiving a unique set of samples.**

## Program design

---

### **Samples - Group 1 Instructions**

Four plastic bottles labelled **PTA 1A, PTA 1B, PTA 2A and PTA 2B** supplied by PTA. The bottles contain approximately 200 mL of artificial potable water.



## Program design

---

### Samples – Group 2 Instructions

Four plastic bottles labelled **PTA 3C**, **PTA 3D**, **PTA 4C** and **PTA 4D** supplied by PTA. The bottles contain approximately 200 mL of artificial potable water.



## Program design

---

### Samples

**PTA 1A** = **PTA 3C**,  
**PTA 1B** = **PTA 3D**,  
**PTA 2A** = **PTA 4C**,  
**PTA 2B** = **PTA 4D**.



## Program design

---

**Programs with small participation the laboratories may be sent samples with unique sample identification.**



## Program design

---

**Lab 1 - Sample 1, Sample 2, Sample 3**

**Lab 2 - Sample A, Sample B, Sample C**

**Lab 3 - Sample 15, Sample 5, Sample 10**

**Lab 4 - Sample Y, Sample Z, Sample X**



## Program design

---

### Identical samples:

**Sample 1**

**Sample A**

**Sample 15**

**Sample Y**

## Action for suspect results

---

**Established programs with many completed rounds may show no evidence of suspect results – so future rounds may need no new design.**

## Action for suspect results

---

- **Accuse participant of dishonest reporting?**
- **Highlight suspect reporting in the commentary in the final report?**
- **Refuse request for future participation?**

## Conclusion

---

- **To identify collusion or falsification of results in a PT program remains a challenge.**
- **The actions following detection of the suspect results are limited and the prevention may be restricted to changes program design.**



---

## **How to contact PTA**

**Proficiency Testing Australia  
PO Box 7507  
Silverwater NSW 2128  
AUSTRALIA**

**Ph: +61 2 9736 8397  
Fax: +61 2 9743 6664  
Email: [ptaenquiry@pta.asn.au](mailto:ptaenquiry@pta.asn.au)  
Web: [www.pta.asn.au](http://www.pta.asn.au)**

**Thank you**